

# URGENT MEDICAL DEVICE RECALL


## AEROWASH Eyewash and Wound Irrigation (15 and 30mL)

Supplied both individually and as part of some Aero branded and custom made first aid kits.

**ARTG: 159438**

Aero Healthcare, following consultation with the Therapeutic Goods Administration, is recalling the below batches of AEROWASH Eyewash and Wound Irrigation due to suspected contamination with *Ralstonia pickettii*.

Eyewash and wound irrigation 15 mL



Eyewash and wound irrigation 30 mL



To see the affected batch numbers please visit

**[aerohealthcare.com.au/recalls](http://aerohealthcare.com.au/recalls).**

Return it to the place of purchase for a refund. The batch number is found on the ampoule label, below the product expiry.

*Ralstonia picketti* is a bacteria commonly found in the natural environment. It rarely causes infection in humans but may cause infection in people with weakened immune systems. Anyone who is concerned in any way about the use of this product should consult their doctor.

Customer Enquiries: [1800 628 881 or [recall@aerohealthcare.com](mailto:recall@aerohealthcare.com)]